# **Pink Spot Vapors**

6485 N. Decatur Blvd. #180 Las Vegas, NV 89131

Carbonyl Assay Samples Received 6/9/15

Product Name Mint Patty

Analysis Report (0615-518)

## GC/MS Analysis For:

Diacetyl 2,3-Pentanedione (aka: Acetyl propionyl)



## Enthalpy Analytical, Inc.

Phone: (919) 850 - 4392 / Fax: (919) 850 - 9012 / www.enthalpy.com 800-1 Capitola Drive Durham, NC 27713-4385



Company:Pink Spot VaporsSample Condition:Samples received in good condition.

Notes: Chain of custody verified.

Method / SOP	Analyte(s)
ENT225	Diacetyl
	2,3-Pentanedione
	(aka: Acetyl propionyl)

#### Summary of Method:

A measured volume of sample is combined with a volume of acetonitrile. A known amount of internal standard (Butanedione-d6) is added and the vial is capped and mixed thoroughly to combine. An aliquot is then analyzed quantitatively against a linear calibration curve using a GC/MS equipped with a mass selective detector. In this analysis, the limit of detection is 1.14 ug/mL for 'Diacetyl' and 1.17 ug/mL for '2,3-Pentanedione'.

Company	Pink Spot Vapors	Cert Report #	0615-556R
Analyst	LAH	# Samples	1
Instrument Type	GC/MS	Sample(s) Received	6/9/15
Project Code	0615-518	Client Project:	na

Enthalpy Code	Sample Name	Brand	Diacetyl (ug/mL)	2,3-Pentanedione (ug/mL)
0615-518-09	Mint Patty	Pink Spot Vapors	1.52 J	Not Detected

### QC Notes:

'2,3-Pentanedione' was not detected above the minimum detection limit.

J: Results flagged "J" are below the Lower Curve Limit, above the MDL, and are considered estimates.

I certify that to the best of my knowledge all analytical data presented in this report:

- Have been checked for completeness
- Are accurate, error-free, and legible
- Have been conducted in accordance with approved protocol, and that all deviations and analytical problems are summarized.

This analytical report was prepared in Portable Document Format (.PDF) and contains 2 pages.

Report Issued: 10/14/2015



Enthalpy Analytical, Inc.

Phone: (919) 850 - 4392 / Fax: (919) 850 - 9012 / www.enthalpy.com 800-1 Capitola Drive Durham, NC 27713-4385